

Notification no. <b>301111874</b>	Notification date <b>29.11.2010</b>	Page: <b>1/1</b>
--------------------------------------	--	---------------------

# Service Report

**Thermo Fisher Scientific**  
**Takkebijsters 1**  
**NL 4817 BL Breda**  
**Phone +31.76.579.55.55**

**Ninovesteenweg 198**  
**B 9320 Erembodegem**  
**service.led.bn1@thermofisher.com**

<b>Location address</b> Universite Catholique de Louvain Laboratoire d' Hépatologie Pédiatrique Tour Vesale +3 Avenue Mounier 52 1200 Bruxelles BE	<b>Contact person:</b> Mme Floriane André  +32 (0)498287869  floriane.andre@uclouvain.be
---	---



Call type <b>Repair Billing</b>	Region <b>200</b>	Field Service Engineer: <b>Desmet, Rudi</b>	
Order date:	by: <b>Floriane André</b>	Customer's order-no. / PO:	Customer cost Center
Device description (Type) <b>LAB.INC.A/J T/C CO2 W/DCN</b>		Contract no.	Contract type:
Serial no. <b>3056932241</b>	Tech. ID no.	Material no. / order no. <b>371</b>	

**Problem description:**  
Check na brand/roetschade (check apres incendie)

**Work done**  
Contrôle apres une incendie  
Incubateur n est pas réparable  
L interieur est completement noire (la suie)  
Toutes les pieces(ventilateur,main pcb,power supply,CO2 valve,transfo,cables,joint porte,les charniers,display pcb,filtrres...)  
a l interieur sont abimés cause la suie  
JOB COMPLETED

Date:	Labour hours:	Travel hours:	Zone:	Comments:
02.12.2010	1.00	3.50	Zone A	

<b>Part requirements</b>			
Part no.	Part description:	Quantity	Quantity Unit

Date / Signature Field Service Engineer		Date / Name / Signature Customer		Work done ?
02.12.2010	 Desmet, Rudi	02.12.2010	 Nyabi Omar	Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>