

Notification no. 301111875	Notification date 29.11.2010	Page: 1/1
--------------------------------------	--	---------------------

Service Report

Thermo Fisher Scientific
Takkebijsters 1
NL 4817 BL Breda
Phone +31.76.579.55.55

Ninovesteenweg 198
B 9320 Erembodegem
service.led.bn1@thermofisher.com

Location address Universite Catholique de Louvain Laboratoire d' Hépatologie Pédiatrique Tour Vesale +3 Avenue Mounier 52 1200 Bruxelles BE	Contact person: Mme Floriane André +32 (0)498287869 floriane.andre@uclouvain.be
---	---



Call type Repair Billing	Region 200	Field Service Engineer: Desmet, Rudi	
Order date:	by: Floriane André	Customer's order-no. / PO:	Customer cost Center
Device description (Type) LAB.INC. 371		Contract no.	Contract type:
Serial no. 3056932244	Tech. ID no.	Material no. / order no. 371	

Problem description:
Check na brand/roetschade (check apres incendie)

Work done
Contrôle apres une incendie
Incubateur n est pas réparable
L interieur est completement noire (la suie)
Toutes les pieces(ventilateur,main pcb,power supply,CO2 valve,transfo,cables,joint porte,les charniers,display pcb,filtrres...) a l interieur sont abimés cause la suie
JOB COMPLETED

Date:	Labour hours:	Travel hours:	Zone:	Comments:
02.12.2010	1.00			

Part requirements			
Part no.	Part description:	Quantity	Quantity Unit

Date / Signature Field Service Engineer	Date / Name / Signature Customer	Work done ?
02.12.2010  Desmet, Rudi	02.12.2010  Nyabi Omar	Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>